

DICOM Correction Proposal

STATUS	Assigned
Date of Last Update	2018/09/13
Person Assigned	Don Van Syckle don.van@dvsconsult.com
Submitter Name	Buck Cunningham buck.cunningham@zeiss.com WG9
Submission Date	2019/09/12

Correction Number	CP-1854
Log Summary:	Change Visual Field Test Point Normals Sequence's Type from 1C to 2C
Name of Standard	PS 3 2018x
Rationale for Correction:	<p>In the current Ophthalmic Visual Field Static Perimetry Measurements IOD, if the Test Point Normals Data Flag is set to YES, a Visual Field Test Point Normals Sequence (Type 1C) is required for EACH test point in the visual field.</p> <p>A normative database may however not provide statistics for all test points in the visual field (a good example is the blind spot where comparison to a normative database doesn't make sense) and it should be possible to skip the Visual Field Test Point Normals Sequence for such points.</p>
Correction Wording:	

In PS 3.3, Section C.8.26.4 Visual Field Static Perimetry Test Measurements Module, Table C.8.26.4-1., make the following changes:

Attribute Name	Tag	Type	Attribute Description
...			
Test Point Normals Data Flag	(0024,0057)	1	Existence of normative data base for this set of test points. Enumerated Values: YES NO
...			
Visual Field Test Point Sequence	(0024,0089)	1	Information for each test point in the visual field. One or more Items shall be included in this Sequence.
...			
>Visual Field Test Point Normals Sequence	(0024,0097)	1C 2C	Information about normal values for each visual field test point.

Attribute Name	Tag	Type	Attribute Description
			<p>One Zero or more Items shall be included in this Sequence.</p> <p>Required if Test Point Normals Data Flag (0024,0057) is YES.</p>