

DICOM Correction Proposal

STATUS	Submitted
Date of Last Update	2017/03/06
Person Assigned	Wim Corbijn
Submitter Name	Wim Corbijn
Submission Date	2017/01/25

Correction Number	CP-1696
Log Summary: Clarify condition for Repetition Time (0018,0080)	
Name of Standard	PS3.3
<p>Rationale for Correction:</p> <p>Current description of the condition is negatively phrased which makes it more difficult to understand. It also has no statement if it can be present otherwise or not.</p> <p>Make the required statement positive and add the statement that the tag “may be present otherwise”.</p>	
Correction Wording:	

Item #1: Update Table C.8-4 Modify Required clause for Repetition Time (0018,0080)

Note

Table C.8-4. MR Image Module Attributes

Attribute Name	Tag	Type	Attribute Description
....			
Repetition Time	(0018,0080)	2C	<p>The period of time in msec between the beginning of a pulse sequence and the beginning of the succeeding (essentially identical) pulse sequence. Required except when Scanning Sequence (0018,0020) is EP and Sequence Variant (0018,0021) is not SK.</p> <p><u>Required if Sequence Variant (0018,0021) is SK or if Scanning Sequence(0018,0020) is not EP.</u></p> <p><u>May be present otherwise.</u></p>
.....			