

DICOM Correction Proposal

STATUS	Submitted
Date of Last Update	2017/05/19
Person Assigned	Wim Corbijn
Submitter Name	Wim Corbijn
Submission Date	2016/11/02

Correction Number	CP-1668
Log Summary:	Add set of Codes for discontinuation of Read Request
Name of Standard	PS3.16
Rationale for Correction:	<p>Add set of codes that can be used to identify the reason for discontinuation of a Read Request. What would be the best way of referencing?</p> <ul style="list-style-type: none"> - Include in CID 9300 - Put Rationale that it is not used in DICOM directly but defined for usage in IHE context. - Leave standalone although not referenced in DICOM but used in IHE
Correction Wording:	

Item #1: Amend PS3.16 Add table with codes that can be used for identifying the reason for the discontinuation of a Read Request.

CID 93XX Read Request Discontinuation Reasons

Type: Extensible

Version: 20xxxxxx

Table CID 93XX. Read Request Discontinuation Reasons

Coding Scheme Designator	Code Value	Code Meaning
DCM	110526	Resource pre-empted
DCM	110500	Doctor canceled procedure
DCM	110502	Incorrect procedure ordered
DCM	110504	Patient died
DCM	110509	Change of procedure for correct charging
DCM	110510	Duplicate order
DCM	110513	Discontinued for unspecified reason

Coding Scheme Designator	Code Value	Code Meaning
DCM	110530	Workitem assignment rejected by assigned resource
DCM	110523	Object Set incomplete