

1	Status	Final Text
2	Date of Last Update	2018/11/10
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	Xavier Carayol
6		mailto:xavier.carayol@csdental.com
7	Submission Date	2018/03/08

8	Correction Number CP-1790	
9	Log Summary: Correct relationship of Reconstruction Algorithm in TID 10013 CT Irradiation Event Data	
10	Name of Standard	
11	PS3.16 2018d	
12	Rationale for Correction:	
13	Only CONTAINER content items may have CONTAINS relationships.	
14	Correction Wording:	

Amend DICOM PS3.16 as follows (changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):

TID 10013 CT Irradiation Event Data

Table TID 10013. CT Irradiation Event Data

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV (113819, DCM, "CT Acquisition")	1	M		
2	>	CONTAINS	TEXT	EV (125203, DCM, "Acquisition Protocol")	1	U		
3	>	CONTAINS	CODE	EV (123014, DCM, "Target Region")	1	M		D???
4	>	CONTAINS	CODE	EV (113820, DCM, "CT Acquisition Type")	1	M		D???
4b	>>	CONTAINS <u>HAS PROPERTIES</u>	CODE	EV (113961, DCM, "Reconstruction Algorithm")	1-n	U		D???
5	>	CONTAINS	CODE	EV (G-C32C, SRT, "Procedure Context")	1	U		D???
...