

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2018/11/11
Person Assigned	Wim Corbijn
Submitter Name	Jeroen Medema
Submission Date	2018/02/22

Correction Number	CP1796
Log Summary: DIMSE Service Element and DIMSE Service Group table caption	
Name of Standard PS3.4 2018d	
Rationale for Correction According to sections 3.5 and 4 of PS3.4, DIMSE means DICOM Message Service Element, so the use of the text "DIMSE Service Element" in PS3.4 is not correct. The tables in PS3.4 that define a DIMSE Service Group do not have consistent captions. It is best to make the contents of the table explicit; then the list of tables at the beginning of the part also makes sense.	
Correction Wording:	

Item #1: Changes in PS3.4, tables F.7.1-1, F.8.1-1, F.9.1-1, H.4-1, H.4-5, H.4.3.1.2-1, H.4.3.2.2-1, H.4.4.2-1, H.4.5.2-1, H.4.6.2-1, H.4.9.2-1, H.4.11.2-1, J.3.1-1, P.2-1, P.3-1, R.3.1-1, S.3.1-1, CC.2-1, CC.2-2, CC.2-3, CC.2-4, DD.3.2-1, EE.2.2-1

DIMSE DICOM Message Service Element	Usage SCU/SCP

Item #2: Changes in PS3.4, section H.3.4

...

- list of supported optional SOP Class Attributes and **DIMSE DICOM Message** Service Elements

...

- list of supported optional SOP Class Attributes and **DIMSE DICOM Message** Service Elements

...

Item #3: Changes in PS3.4, section EE.3.1

...

- list of supported SOP Class Attributes and **DIMSE DICOM Message** Service Elements

...

- list of supported SOP Class Attributes and **DIMSE DICOM Message** Service Elements

...

Item #4: Changes in PS3.4

...

Table F.7.1-1. DIMSE Service Group Applicable to Modality Performed Procedure Step

...

Table F.8.1-1. DIMSE Service Group Applicable to Modality Performed Procedure Step Retrieve

...

Table F.9.1-1. DIMSE-N Service Group Applicable to Modality Performed Procedure Step Notification

...

Table H.4-1. DIMSE Service Group Applicable to Basic Film Session

...

Table H.4-5. DIMSE Service Group Applicable to Basic Film Box

...

Table H.4.3.1.2-1. DIMSE Services Group Applicable to Basic Grayscale Image Box

...

Table H.4.3.2.2-1. DIMSE Services Group Applicable to Color Grayscale Image Box

...

Table H.4.4.2-1. DIMSE Services Group Applicable to Basic Annotation Box

...

Table H.4.5.2-1. DIMSE Services Group Applicable to Print Job

...

Table H.4.6.2-1. DIMSE Services Group Applicable to Printer

...

Table H.4.9.2-1. DIMSE Services are Group Applicable to Presentation LUT

...

Table H.4.11.2-1. ~~ID~~-DIMSE Services Group Applicable to Printer Configuration Retrieval

...

Table J.3.1-1. ~~ID~~-DIMSE Services Group Applicable to Storage Commitment Push Model

...

Table P.2-1. DIMSE Service Group Applicable to Procedural Event Logging

...

Table P.3-1. DIMSE Service Group Applicable to Substance Administration Logging

...

Table R.3.1-1. DIMSE Service Group Applicable to Instance Availability Notification

...

Table S.3.1-1. DIMSE-N Services Group Applicable to Media Creation Management

...

Remove the dashes in the following figure titles:

Table CC.2-1. DIMSE Service Group ~~Applicable to~~ UPS Push

...

Table CC.2-2. DIMSE Service Group ~~Applicable to~~ UPS Pull

...

Table CC.2-3. DIMSE Service Group –Applicable to UPS Watch

...

Table CC.2-4. DIMSE Service Group –Applicable to UPS Event

...

Table DD.3.2-1. DIMSE Service Group Applicable to Machine Verification

...

Table EE.2.2-1. DIMSE Service Group –Applicable to Display System