

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2018/11/11
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Sven Siekmann (sven.siekmann@brainlab.com) Christof Schadt (christof.schadt@brainlab.com)
Submission Date	2018/06/18

Correction Number	cp1840
Log Summary:	Make Code Meaning consistent
Name of Standard	PS 3.16
Rationale for Correction:	In Part 16 there are two codes describing a laterality of “Bilateral” (CID 2) or “Left and right” (CID 244). According to SNOMED both Code Meanings are possible. Still, it might be worth to adjust the code that was introduced later in CID 244 to “Bilateral”.
Correction Wording:	

Change in Part 16, CID 244

CID 244 Laterality

Resources: [HTML](#) | [FHIR JSON](#) | [FHIR XML](#) | [IHE SVS XML](#)
 Type: Non-Extensible
 Version: 20030108
 UID: 1.2.840.10008.6.1.37

Table CID 244. Laterality

Coding Scheme Designator	Code Value	Code Meaning	SNOMED-CT Concept ID	UMLS Concept Unique ID
SRT	G-A100	Right	24028007	C0205090
SRT	G-A101	Left	7771000	C0205091
SRT	G-A102	Right and left <u>Bilateral</u>	51440002	C0238767
SRT	G-A103	Unilateral	66459002	C0205092