

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2016/01/19
Person Assigned	Harry Solomon
Submitter Name	Renate Höcker <a href="mailto:Renate.Hoecker@siemens.com">Renate.Hoecker@siemens.com</a>
Submission Date	2015/02/18

Correction Number	CP-1462
Log Summary: Admission ID missing in MPPS module	
Name of Standard	PS3.3 2015c
Rationale for Correction: Admission ID and Issuer of Admission ID Sequence are missing in Performed Procedure Step Relationship Module. They are only mentioned in part 4 Table F.7.2-1.	
Correction Wording:	

**Table C.4-13. Performed Procedure Step Relationship Module Attributes**

Attribute Name	Tag	Attribute Description
...		
Referenced Patient Sequence	(0008,1120)	Uniquely identifies the Patient SOP Instance. Only a single Item shall be included in this sequence.
<i>&gt;Include Table 10-11 "SOP Instance Reference Macro Attributes"</i>		
<b><u>Admission ID</u></b>	<b><u>(0038,0010)</u></b>	<b><u>Identification number of the visit as assigned by the healthcare provider</u></b>
<b><u>Issuer of Admission ID Sequence</u></b>	<b><u>(0038,0014)</u></b>	<b><u>Identifier of the Assigning Authority that issued the Admission ID.</u></b> <b><u>Only a single Item shall be included in this sequence.</u></b>
<i>&gt;Include Table 10-17 "HL7v2 Hierarchic Designator Macro Attributes"</i>		
Service Episode ID	(0038,0060)	Identifier of the Service Episode as assigned by the healthcare provider
...		