

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2017/04/13
Person Assigned	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submitter Name	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submission Date	2016/08/30

Correction Number	CP-1661
Log Summary:	Correct Code in CID 9250
Name of Standard	PS3.16 2017a
Rationale for Correction:	CID 9250 Scheduled Processing Parameter Concept Codes for RT Treatment contains an inconsistent pair of Code Value and Code Meaning: It references Code Value '121712' and Code Meaning 'Treatment Delivery Type', which are inconsistent. The Code Value needs to be '121740'.
Correction Wording:	

*In PS 3.16, correct CID 9250 as follows*

### **CID 9250 Scheduled Processing Parameter Concept Codes for RT Treatment**

Type: Extensible  
Version: 2014040220170413

**Table CID 9250. Scheduled Processing Parameter Concept Codes for RT Treatment**

Coding Scheme Designator	Code Value	Code Meaning
DCM	<del>121712</del> <u>121740</u>	Treatment Delivery Type