

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2017/04/18
Person Assigned	Don Van Syckle (don.van@dvsconsult.com)
Submitter Name	Don Van Syckle (don.van@dvsconsult.com) for WG 9
Submission Date	2016/09/07

Correction Number	CP-1662
Log Summary: Define the encoding order for OPT Attribute Reference Coordinates (0022,0032)	
Name of Standard PS3.3 2017a	
<p>Rationale for Correction:</p> <p>Attribute Reference Coordinates (0022,0032) is used in the OPT IOD so a frame can be described in terms of its position (row/column coordinates) to another image (i.e., typically an OP image used as a scout image for the OPT slice). The order of encoding of the coordinates is not explicitly defined in the attribute description.</p> <p>Explanatory text about the attribute implies the order as row\column (see below) and also WG9 has investigated this with multiple (4) OPT device implementations and all implementations encode the data as row/column. So the attribute description needs to be updated to convey this choice.</p> <p>The “implied wording” already in the document in Section C.8.17.10.1.1 is as follows:</p> <p>“b. Reference Coordinates (0022,0032) would contain values that give the row and column coordinates for the image pixel on the referenced image that corresponds to the first column of the OPT Frame 1 and the row and column coordinates that corresponds to the last column of the OPT Frame 1.”</p>	
Correction Wording:	
Update attribute Reference Coordinates (0022,0032) in Table C.8.17.10-1	

Table C.8.17.10-1. Ophthalmic Frame Location Macro Attributes

Attribute Name	Tag	Type	Attribute Description
Ophthalmic Frame Location Sequence	(0022,0031)	1	Specifies the column locations for this frame in terms of locations on a referenced image. One or more Items shall be included in this Sequence.
<i>>Include Table 10-3 “Image SOP Instance Reference Macro Attributes”</i>			
>Purpose of Reference Code Sequence	(0040,A170)	3	Describes the purpose for which the reference is made. Only a single Item is permitted in this Sequence.
<i>>>Include Table 8.8-1 “Code Sequence Macro Attributes”</i>			
>Reference Coordinates	(0022,0032)	1	Defined Code Concept for Purpose of Reference is (121311, DCM, "Localizer") Image coordinates for the points on the referenced image that correspond to the points on this frame. See Section C.8.17.10.1.1. Multiple pairs of values where the first value of

Attribute Name	Tag	Type	Attribute Description
			<p>each pair is the row and the second value of each pair is the column. Row is the vertical offset and column is the horizontal offset.</p> <p>Specified with sub-pixel resolution such that the origin at the Top Left Hand Corner (TLHC) of the TLHC pixel is 0.0\0.0, the Bottom Right Hand Corner (BRHC) of the TLHC pixel is 1.0\1.0, and the BRHC of the BRHC pixel is Rows\Columns (see figure C.10.5-1, except that that row and column order is reversed). The values must be within the range 0\0 to Rows\Columns.</p>
>.....