

1	Status	Letter Ballot
2	Date of Last Update	2017/04/04
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	David Clunie
6		mailto:dclunie@dclunie.com
7	Submission Date	2016/11/21

8	Correction Number CP-1673	
9	Log Summary: Remove Version-specific UIDs from Confidentiality Profile Attributes	
10	Name of Standard	
11	PS3.15	
12	Rationale for Correction:	
13	Creator Version UID and Context Group Extension Creator UID are not "transient" UIDs that leak identifying information and which	
14	need to be removed or remapped, but rather are important for understanding the meaning of the scope that they defines, e.g., private	
15	raw data elements in Raw Data IODs.	
16	Correction Wording:	

1  
2

Amend DICOM PS3.15 as follows (changes to existing text are bold and ~~strikethroughd~~ for additions and ~~struckthrough~~ for removals):

3

**Table E.1-1. Application Level Confidentiality Profile Attributes**

4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15

Attribute Name	Tag	Retired (from ????)	In Std. Comp. IOD (from ????)	Basic Profile	Retain Safe Private Option	Retain UIDs Option	Retain Device Ident. Option	Retain Patient Chars. Option	Retain Long. Full Dates Option	Retain Long. Modif. Dates Option	Clean Desc. Option	Clean Struct. Cont. Option	Clean Graph. Option
<del>Context Group Extension Creator UID</del>	<del>(0008,010B)</del>	<b>N</b>	<b>Y</b>	<b>U</b>		<b>K</b>							
<del>Creator Version UID</del>	<del>(0008,9123)</del>	<b>N</b>	<b>Y</b>	<b>U</b>		<b>K</b>							