

1	Status	Assigned
2	Date of Last Update	2017/01/24
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	David Clunie
6		mailto:dclunie@dclunie.com
7	Submission Date	2016/12/07

8	Correction Number CP-1677	
9	Log Summary: Do not include Modifying Device Manufacturer in Confidentiality Profile Attributes	
10	Name of Standard	
11	PS3.15	
12	Rationale for Correction:	
13	Modifying Device Manufacturer (0020,3404) should not have been included in the list of Attributes to be removed (Manufacturer	
14	(0008,0070) is not).	
15	Correction Wording:	

Amend DICOM PS3.15 as follows (changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):

Table E.1-1. Application Level Confidentiality Profile Attributes

Attribute Name	Tag	Retired (from ????)	In Std. Comp. IOD (from ????)	Basic Profile	Retain Safe Private Option	Retain UIDs Option	Retain Device Ident. Option	Retain Patient Chars. Option	Retain Long. Full Dates Option	Retain Long. Modif. Dates Option	Clean Desc. Option	Clean Struct. Cont. Option	Clean Graph. Option
Modifying Device Manufacturer	(0020,3404)	Y	N	X									