

DICOM Correction Proposal

STATUS	Submitted
Date of Last Update	2017/03/06
Person Assigned	Wim Corbijn
Submitter Name	Wim Corbijn
Submission Date	2017/01/20

Correction Number	CP-1688
Log Summary: Correct Note of Table C.7.6.18-1 Cardiac Synchronization Module Attributes	
Name of Standard PS3.3	
Rationale for Correction: Note of Table C.7.6.18-1 Cardiac Synchronization Module Attributes is stating "The Cardiac Gating Sequence (0018,9118) in the Cardiac Gating Functional Group uses the same attributes and specifies the values for a single frame." Whereas (0018,9118) is called Cardiac Synchronization Sequence and the Functional group is called Cardiac Synchronization Functional Group. The Note shall use the correct names.	
Correction Wording:	

Item #1: Correct PS3.3 note under Table C.7.6.18-1 to use the correct Attribute and Functional Group names.

Note

Low R-R Value (0018,1081), High R-R Value (0018,1082), Intervals Acquired (0018,1083) and Intervals Rejected (0018,1084) in this Module apply to the entire Image. The Cardiac **Gating Synchronization** Sequence (0018,9118) in the Cardiac **Gating Synchronization** Functional Group uses the same attributes and specifies the values for a single frame.