

1	Status	Final Text
2	Date of Last Update	2016/11/05
3	Person Assigned	David Clunie
4		mailto:dclunie@dclunie.com
5	Submitter Name	Jörg Riesmeier
6		mailto:mail@jriesmeier.com
7	Submission Date	2015/08/21

8	Correction Number CP-1578	
9	Log Summary: Correct Pulse Sequence Name in MR Image Library	
10	Name of Standard	
11	PS3.16 2016d	
12	Rationale for Correction:	
13	The code used for Pulse Sequence Name in TID 1606 is already used for something else.	
14	The definition in PS3.3 for the corresponding attribute is used.	
15	Correction Wording:	

Amend DICOM PS3.16 as follows (changes to existing text are bold and underlined for additions and ~~struckthrough~~ for removals):

TID 1606 Image Library Entry Descriptors for MR

Table TID 1606. Image Library Entry Descriptors for MR

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1		HAS ACQ CONTEXT	TEXT	EV (110909 <u>128230</u> , DCM, "Pulse Sequence Name")	1	U		

Content Item Descriptions

Pulse Sequence Name	The value of Pulse Sequence Name (0018,9005) or Sequence Name (0018,0024) in the Image IOD. See PS3.3 Section C.8.13.4 "MR Pulse Sequence Module".
---------------------	--

D DICOM Controlled Terminology Definitions (Normative)

This Annex specifies the meanings of codes defined in DICOM, either explicitly or by reference to another part of DICOM or an external reference document or standard.

DICOM Code Definitions (Coding Scheme Designator "DCM" Coding Scheme Version "01")

Table D-1. DICOM Controlled Terminology Definitions

Code Value	Code Meaning	Definition	Notes
110909	Image Orientation (Patient) Column Z	The z value of the first column direction cosine with respect to the patient, with respect to the patient-based coordinate system.	
<u>128230</u>	<u>Pulse Sequence Name</u>	<u>Name of an MR pulse sequence for annotation purposes. Potentially vendor-specific name.</u>	