

DICOM Correction Proposal

STATUS	Assigned
Date of Last Update	2016/08/30
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Ulrich Busch (ulrich.busch@varian.com)
Submission Date	2016/08/30

Correction Number	CP-1661
Log Summary:	Correct Code in CID 9250
Name of Standard	PS 3.16 2016c
Rationale for Correction:	CID 9250 Scheduled Processing Parameter Concept Codes for RT Treatment contains an inconsistent pair of Code Value and Code Meaning: It references Code Value '121712' and Code Meaning 'Treatment Delivery Type', which are inconsistent. The Code Value needs to be '121740'.
Correction Wording:	

In PS 3.16, correct CID 9250 as follows

CID 9250 Scheduled Processing Parameter Concept Codes for RT Treatment

Version: Type: Extensible 20140402

Table CID 9250. Scheduled Processing Parameter Concept Codes for RT Treatment

Coding Scheme Designator	Code Value	Code Meaning
DCM	121712 <u>121740</u>	Treatment Delivery Type