

## DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2014/09/08
Person Assigned	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submitter Name	Christof Schadt ( <a href="mailto:christof.schadt@brainlab.com">christof.schadt@brainlab.com</a> )
Submission Date	2014/03/20

Correction Number	CP-1398
Log Summary: Add FOR Module to RTSS IOD	
Name of Standard PS 3 2011	
Rationale for Correction:  By declaring the Frame of Reference Relationship Sequence (3006,00C0) as deprecated with CP 1249, there is no reason that the Frame of Reference Module is not part of the RT Structure Set Information Object Definition.  As it also eases the interpretation of this object by high-level DICOM interpreters, the proposal is to add this Module to the IOD.	
Correction Wording:  <include proposed change below, following guidelines for formatting of changes in supplements>	

*Adapt A.19.3 RT Structure Set IOD Module Table*

**Table A.19.3-1—RT STRUCTURE SET IOD MODULES**

IE	Module	Reference	Usage
Patient	Patient	C.7.1.1	M
	Clinical Trial Subject	C.7.1.3	U
Study	General Study	C.7.2.1	M
	Patient Study	C.7.2.2	U
	Clinical Trial Study	C.7.2.3	U
Series	RT Series	C.8.8.1	M
	Clinical Trial Series	C.7.3.2	U
<b><u>Frame of Reference</u></b>	<b><u>Frame of Reference</u></b>	<b><u>C.7.4.1</u></b>	<b><u>U – See Note 2</u></b>
Equipment	General Equipment	C.7.5.1	M
Structure Set	Structure Set	C.8.8.5	M
	ROI Contour	C.8.8.6	M
	RT ROI Observations	C.8.8.8	M
	Approval	C.8.8.16	U
	SOP Common	C.12.1	M

Note: 1. The Audio Module was previously included in this IOD but has been retired. See PS 3.3 2004.  
2. Previously the Frame of Reference Module was not included (see PS 3.3 2011). In case more than one Frame of Reference is referenced within the Structure Set Module's Referenced Frame of Reference Sequence (3006,0010) the Frame of Reference Module will not be present.