

DICOM Correction Item

Correction Number		CP 754
Log Summary: Add Display Shutter Module to CR IOD		
Type of Modification	Name of Standard	
Addition	PS 3.3 2007	
Rationale for Correction		
The Display Shutter is useful for CR applications, and often added by vendors anyway.		
Sections of documents affected		
PS 3.3 A.2.3		
Correction Wording:		

Add the Display Shutter Module to overview table and CR Image IOD Module Table:

A.2.3 CR Image IOD Module Table

**Table A.2-1
 CR IMAGE IOD MODULES**

IE	Module	Reference	Usage
...
Image	General Image	C.7.6.1	M
	Image Pixel	C.7.6.3	M
	Contrast/bolus	C.7.6.4	C - Required if contrast media was used in this image
	<u>Display Shutter</u>	<u>C.7.6.11</u>	<u>U</u>
	Device	C.7.6.12	U
...