

DICOM Correction Proposal Form

Tracking Information - Administration Use Only	
Correction Proposal Number	CP-273
STATUS	Proposed
Date of Last Update	2001/06/20
Person Assigned	Charles Parisot charles.parisot@med.ge.com
Submitter Name	Charles Parisot charles.parisot@med.ge.com
Submission date	2001/06/20

Correction Number	CP-273
Log Summary: Identical Document Sequence Typo	
Type of Modification	Name of Standard
Correct Typo	PS 3.3 2000
Rationale for Correction	
The Table C.17.6-2 KEY OBJECT DOCUMENT MODULE ATTRIBUTES entry on Identical Document Sequence is inconsistent with the text in Section C.17.6.2.1. In the table it states incorrectly "one or more studies" when it should be "more than one study".	
Sections of documents affected	
Table C.17.6-2 PS 3.3	
Correction Wording:	

Identical Documents Sequence	(0040,A525)	1C	Duplicates of this document, stored with different SOP Instance UIDs. One or more Items may be included in this sequence. Required if this Key Object Selection document references instances in one or more than one other Studies. See C.17.2.2 and C.17.6.2.1 for further explanation and conditions.
------------------------------	-------------	----	--

C.17.6.2.1 Identical Documents

If the Current Requested Procedure Evidence Sequence (0040,A375) references SOP Instances both in the current study and in one or more other studies, this document shall be duplicated into each of those other studies, and the duplicates shall be referenced in the Identical Documents Sequence (0040,A525).

Note: Thus a Key Object Selection Document that references images in the current study as well as in a prior or comparison study, would be duplicated into the other study. This allows an application displaying that other study to easily access notes relevant to that study's SOP Instances.