



Secretariat: MITA/NEMA
1300 North 17th Street, Suite 900
Arlington, VA 22209, USA
+1-703-841-3259
<http://www.dicomstandard.org>
dicom@dicomstandard.org

Meeting Minutes

DICOM WORKING GROUP SEVEN (RADIOTHERAPY)

December 17, 2024

Participants:

Jim Percy (AAPM Affiliate) and DICOM WG 7 Co-Chair
Christof Schadt (Brainlab) and DICOM WG 7 Co-Chair Rollo Moore
Kari Jyrkkala (Varian)
David Holst (RaySearch)
Emily Hofer (NEMA)
Marcus Bergman (RaySearch)
Walter Bosch (AAPM)
Uli Busch (personal)
Wilhelm Sjøsted (RaySearch)
Yulong Yang (AAPM)
Bruce Curran (AAPM)
Michael Owens (Reflexion)

Meeting came to order at 11:00am EST

The group reviewed the following Correction Proposals:

cp_RT261 Alternate Machine Module for RTPlan (version last updated 10/18/2024)

- RT Beams Module Attributes
 - Review usage of existing RT Plan Attributes for alternate machine plans
 - Add note at the beam level to indicate that beam parameters are “representative” values and that actual / detailed values are contained in the RT Alternate Machine Data Module. The “representative” values are not used for plan delivery or dose calculation. This note is directed primarily to plan consumers (TMS, QA systems, etc.).
- Decision to include Alternate Machine Data in the same format for treatment record.
- RT Alternate Machine Data
- Jim to clean up Alternate Machine Data Module and start working on treatment record

- Next session on Jan 7, 2025

Meeting adjourned at 11:48am EST