

## DICOM Correction Proposal Form

Tracking Information - Administration Use Only	
Correction Proposal Number	CP-356
STATUS	Voting Packet
Date of Last Update	3/27/2003
Person Assigned	Harry Solomon
Submitter Name	Harry Solomon
Submission date	11/22/2002

Correction Number	CP-356
Log Summary: Correct Acquisition Context Module note in DX IOD	
Type of Modification	Name of Standard
Clarification	PS 3.3
<p>Rationale for Correction:</p> <p>The Acquisition Context Module has a Type 2 sequence attribute, so it is never empty (no attribute), but the sequence may be empty.</p>	
<p>Sections of documents affected</p> <p>PS 3.3 Section A.26.3</p>	
<p>Correction Wording:</p> <p>Notes: ...</p> <p>4. The Acquisition Context Module is mandatory, but may <del>be</del> <b>include only an empty (zero-length) Acquisition Context Sequence (0040,0555)</b> <del>The intent is that</del> <b>Thus</b> all <b>Level 1 or 2</b> Storage SCPs will preserve any information present, <b>without requiring and</b> acquisition systems <del>to be</del> <b>are not</b> required to generate any contents <del>in that sequence, or requiring display systems to use the information for annotation (which is beyond the scope of DICOM).</del></p>	