

DICOM Correction Item

Correction Number CP-451	
Log Summary: Add Key Object Selection Document Title for Report Attachment	
Type of Modification Modification	Name of Standard PS 3.16
<p>Rationale for Correction:</p> <p>A major major increment in radiology reporting functionality can be achieved by a simple, automated process that can attach key images to a clinical report. The problem today is that transcription of reports comes through a system path disjoint from the image review path. A Key Object Selection Document, created at the image review station and identifying images to be appended to a report, can be used to pass image references to the report management system; those images can then be appended to the report by an automated process after transcription is complete.</p> <p>This workflow has been identified by WG-10 as a high priority based on their 2003 analysis of reporting workflows. (See presentation given at DICOM Anniversary Conference.)</p>	
<p>Sections of documents affected</p> <p>PS 3.16 Annex B and Annex D</p>	
Correction Wording:	

Annex B

CID 7010 Key Object Selection Document Title

Context ID 7010

Key Object Selection Document Title

Type: Extensible Version: 200209042004mmdd

Coding Scheme Designator (0008,0102)	Code Value (0008,0100)	Code Meaning (0008,0104)
...		
<u>99CPxxx</u>	<u>x113020</u>	<u>For Report Attachment</u>

Annex D

DICOM Code Definitions (Coding Scheme Designator “DCM” Coding Scheme Version “01”)

Code Value	Code Meaning	Definition	Notes
...			
<u>x113020</u>	<u>For Report Attachment</u>	<u>Selection of information objects for attachment to the clinical report of the Current Request Procedure</u>	