

## DICOM Correction Item

Correction Number	CP-580
Log Summary: Correct abstract prior codes for hanging protocols	
Type of Modification	Name of Standard
Clarification	PS 3.16 2004
Rationale for Correction	
The codes used in Sup 60 were already in use for another purpose in the 2004 standard.	
Sections of documents affected	
PS 3.16 Annex B and D	
Correction Wording:	

*Item #1: Amend the following context group in PS 3.16, Annex B:*

### **CID 31            Abstract Priors**

This Context Group includes codes that may be used to identify imaging procedures that may be referred to as priors for the purpose of image set selection in Hanging Protocols.

#### **Context ID 31 Abstract Priors**

**Type: Extensible            Version: 20050111yyyymmdd**

Coding Scheme Designator	Code Value	Code Meaning
SRT	G-7292	On admission
SRT	R-400B2	Intraoperative
SRT	R-41FD9	Pre-admission
SRT	R-411C0	Pre-dose
SRT	R-404DA	Post-dose
SRT	R-413C5	Pre-operative
SRT	R-413B7	Post-operative
DCM	<del>109110</del> <u>109120</u>	On admission to unit
DCM	<del>109111</del> <u>109121</u>	On discharge
DCM	<del>109112</del> <u>109122</u>	On discharge from unit
DCM	<del>109113</del> <u>109123</u>	Pre-intervention
DCM	<del>109114</del> <u>109124</u>	Post-intervention
DCM	<del>109115</del> <u>109125</u>	At last appointment

*Item #2: Amend the following definitions to PS 3.16, Annex D:*

109110	Voice	The sound of a human's speech, recorded during a procedure.	May include the patient's voice, or the voice of staff present in the room, or an operator's voice (whether for the purpose of recording a narrative accompanying a procedure or not).
109111	Operator's narrative	The voice of a device operator, recorded during a procedure.	
109112	Ambient room environment	The ambient sound recorded during a procedure, which may or may not include voice and other types of sound.	
109113	Doppler audio	The Doppler waveform recorded as an audible signal.	
109114	Phonocardiogram	The sound of the human heart beating.	Such as might be recorded from an electronic stethoscope.
109115	Physiological audio signal	Any sound made by the human body.	May include the sound of the heart, but also sound from other organs, such as bowel sounds or bruits from vessels, or sounds of respiration. Not intended to include voice.
<del>109110</del> <u>109120</u>	On admission to unit	The occasion on which a procedure was performed on admission to a specialist unit (e.g., intensive care).	
<del>109111</del> <u>109121</u>	On discharge	The occasion on which a procedure was performed on discharge from hospital as an in-patient.	
<del>109112</del> <u>109122</u>	On discharge from unit	The occasion on which a procedure was performed on discharge from a specialist unit (e.g., intensive care).	
<del>109113</del> <u>109123</u>	Pre-intervention	The occasion on which a procedure was performed immediately prior to non-surgical intervention (e.g, percutaneous angioplasty, biopsy).	
<del>109114</del> <u>109124</u>	Post-intervention	The occasion on which a procedure was performed immediately after to non-surgical intervention (e.g, percutaneous angioplasty, biopsy).	
<del>109115</del> <u>109125</u>	At last appointment	The occasion on which a procedure was performed at the most recent outpatient visit.	

CP-580  
Correct abstract prior codes for hanging protocols

Date: 2005/11/02  
Status: Letter Ballot