

DICOM Correction Item

Correction Number	CP-653
Log Summary: Clarify SR Unambiguous Rendering Requirement	
Type of Modification	Name of Standard
Clarification	PS 3.4
Rationale for Correction	
The unambiguous rendering requirement for SR objects is not limited to just the content tree, but applies to the whole SOP Instance. A note is added to clarify this requirement.	
Sections of documents affected	
PS 3.4 Section O.2.2	
Correction Wording:	

O.2.2 Behavior of an SCP

An SCP intending to display or otherwise render a Structured Report shall convey its full meaning in an unambiguous manner.

Note: “Full meaning” includes not just the Content Tree (i.e., the Items of the Content Sequence), but all attributes of the Data Set that are necessary to properly interpret the Structured Report. This includes those attributes that set the initial Observation Context for the Content Tree, i.e., the patient, procedure, and observer identifiers, and the Completion status and Verification status of the Structured Report.