

DICOM Correction Item

Correction Number CP 716	
Log Summary: Stereometric Relationship Series Information Missing	
Type of Modification Addition	Name of Standard PS 3.3 2007
Rationale for Correction: The Stereometric Relationships IOD is missing the general series information. The entity-relationship model indicates that this IOD is part of a series. The information in the General Series Module is required to properly identify the series' to which a specific instance belongs.	
Sections of documents affected PS 3.3 Tables A.1-2, A.43-2	
Correction Wording:	

Add General Series as M to Stereometric Relationship IOD in Table A.1-2

Add General Series as M to Stereometric Relationship IOD in Table A.43-2:

A.43.2 Stereometric Relationship IOD Modules

**Table A.43-2
STEREOMETRIC RELATIONSHIP IOD MODULES**

IE	Module	Reference	Usage
Patient	Patient	C.7.1.1	M
	Specimen Identification	C.7.1.2	C - Required if the Imaging Subject is a Specimen
	Clinical Trial Subject	C.7.1.3	U
Study	General Study	C.7.2.1	M
	Patient Study	C.7.2.2	U
	Clinical Trial Study	C.7.2.3	U
Series	General Series	C.7.3.1	M
	Clinical Trial Series	C.7.3.2	U
	Stereometric Series	C.8.18.1	M
...