

DICOM Correction Item

Log Summary: Correct Results Interpretation Relationship Module	
Type of Modification Correction	
Name of Document PS 3.3	Version Number 1999
Rationale for change In the 1996 version of the standard, the relationship in the DICOM Information Model between Results IOD and Interpretation IOD is 0-1 (see PS3.3 Figure 7-2). In the Results Relationship Module (PS3.3 Table C.5-1), the Referenced Interpretation Sequence seems to imply multiple items are allowed, but this conflicts Figure 7-2. In the 1998 version of the standard, the Referenced Interpretation Sequence is clarified to specify zero or more items are allowed. The figure, now 7-2a, is not changed. Figure 7-2a should be changed to specify the relationship between Results IOD and Interpretation IOD to 0-n	
Sections of document affected PS3.3 Figure 7-2a	

