

DICOM Correction Proposal Form

Tracking Information - Administration Use Only	
Correction Proposal Number	CP-nnn
STATUS	Assigned
Date of Last Update	2000/09/27
Person Assigned	David Clunie dclunie@idt.net
Submitter Name	Janet Keyes
Submission date	2000/09/14

Correction Number	CP-xxx
Log Summary: Restrict SR COMPOSITE, IMAGE and WAVEFORM to one reference	
Type of Modification	Name of Standard
Clarification	PS 3.3 2000
Rationale for Correction Unlike all the other SR value types which allow only a single value, the COMPOSITE type allows 1 or more objects (and hence so does IMAGE and WAVEFORM). The singular in the introductory text conflicts with the plural in the macro table. Should not this be restricted to one reference (probably a cut and paste error from the use of the referenced SOP class/instance UID text from elsewhere) ?	
Sections of documents affected PS 3.3 C.18.3	
Correction Wording:	

C.18.3 Composite Object Reference Macro

Table C.18.3-1 specifies the Attributes that convey a reference to a DICOM Composite Object that is not a DICOM Image or Waveform (such as an SR Document).

- Notes: 1. If a Softcopy Presentation State is to be applied to an Image, it should be referenced by an Image Reference Macro.
 2. Other SR Documents may be referenced by this macro, but there is no facility to reference individual Content Items within those reports.

**Table C.18.3-1
 COMPOSITE OBJECT REFERENCE MACRO ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
Referenced SOP Sequence	(0008,1199)	1	References to Composite Object SOP Class/SOP Instance pairs. One or more items may be included in this sequence. Only a single item shall be permitted in this sequence.
>Referenced SOP Class UID	(0008,1150)	1	Uniquely identifies the referenced SOP Class..

CP-216
Restrict SR COMPOSITE to one reference

Date 2000/09/29
Status: Assigned

>Referenced SOP Instance UID	(0008,1155)	1	Uniquely identifies the referenced SOP Instance.
------------------------------	-------------	---	--