

2

DICOM Correction Proposal Form

Tracking Information - Administration Use Only	
Correction Proposal Number	CP-227
STATUS	Assigned
Date of Last Update	2000/11/01
Person Assigned	Andrew Hewett andrew.hewett@med.siemens.de
Submitter Name	Andrew Hewett andrew.hewett@med.siemens.de Heinz Blendinger heinz.blendinger@med.siemens.de
Submission date	2000/10/20

4

Correction Number	CP-227
Log Summary: Add Comments on the Performed Procedure Step to General Series Module	
Type of Modification	Name of Standard
Addition	PS 3.3 2000
<p>Rationale for Correction</p> <p>The attribute Comments on the Performed Procedure Step (0040,0280) is an element of the Performed Procedure Step Information Module.</p> <p>The General Series Module (Table C-7.4) contains some attributes of the Performed Procedure Step but does not include Comments on the Performed Procedure Step (0040,0280). If included, this attribute would provide the possibility to include such comments within SOP Instances.</p> <p>Note that the definition of the existing attribute Performed Procedure Step Description (0040,0254) is intended for a fixed description or classification but does not provide the required "user comment" capability.</p>	
Sections of documents affected	
PS 3.3 C.7.3.1	
Correction Wording:	

CP-277
Add Comments on the Performed Procedure Step
to General Series Module

Date 2000/10/06

Status: Assigned

2

Item 1: Add the following lines to Table C-7.4 in PS 3.3

4

**Table C.7-4
GENERAL SERIES MODULE ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
...			
Comments on the Performed Procedure Step	(0040,0280)	3	User-defined comments on the Performed Procedure Step.
...			

6