

DICOM Correction Item

Correction Number: CP-1088	
Log Summary: Modality Code for Structured Display	
Type of Modification	Name of Standard PS 3.3, PS 3.16 – 2009
Rationale for Correction The Structured Display IOD makes use of General Series Module and does not specify / restrict the value for Modality Code (0008,0060). Since the IOD has been in the Standard for some time it is proposed to add a new value to	
Sections of documents affected PS 3.3, Section A.33.5, C.7.3.1.1.1 ; PS 3.16 Annex D	
Correction Wording: n.a.	

Item 1: Add Modality Code Constraint to PS 3.3, A.33.5

A.33.5.2 Basic Structured Display IOD Content Constraints

A.33.5.2.1 Modality

The value of Modality (0008,0060) shall be SD.

Item 2: Add new Modality Code to General Series Module PS 3.3, C.7.3.1.1.1

C.7.3.1.1.1 Modality

Defined Terms for the Modality (0008,0060) are:

SD = Structured Display

Item 3: Add to PS3.16 – Add To Annex D DICOM Controlled Terminology Definitions (Normative)

DICOM Code Definitions (Coding Scheme Designator “DCM” Coding Scheme Version “01”)

Code Value	Code Meaning	Definition	Notes
...
<u>SD</u>	<u>Structured Display</u>	<u>Structured Display Object</u>	
...