

DICOM Change Proposal Form - Final Text

Tracking Information - Administrator Use Only	
Change Proposal Number	CP-21
STATUS	voting-1, approved
Date of Last Update	9/14/94
Person Assigned	Fred Prior

Log Summary: Same Attribute in multiple modules	
Submitter Name F.Prior	Submission date 9/14/94
Type of Modification Clarification	
Name of Document PS 3.3	Version Number
<p>Rationale for change</p> <p>When multiple modules containing the same Attribute(s) are included in an IOD, the Attribute should only included once in the data set. This rule is not included in the standard.</p>	
<p>Sections of document affected</p> <p>6.2</p>	
<p>Suggested Wording of change</p> <p>Add after the second Paragraph:</p> <p>When multiple modules containing the same Attribute(s) are included in an IOD, the Attribute shall be encoded only once into a Data Element. If one of the modules is a specialization of the other (e.g., General Image and CT Image) the Attribute definition contained in the more specialized module shall be used.</p>	