

DICOM Change Proposal Form - Final Text

Tracking Information - Administrator Use Only	
Change Proposal Number	CP-25
STATUS	voting-1, approved
Date of Last Update	9/15/94
Person Assigned	Fred Prior

Log Summary: Frame of Reference note - relating CT/MR slices to localizer	
Submitter Name F.Prior	Submission date 9/14/94
Type of Modification Clarification	
Name of Document PS 3.3	Version Number
Rationale for change Relating CT and MR slices to Localizer images.	
Sections of document affected C.7.4.1.1.1	
Suggested Wording of change Add a second note: A common Frame of Reference UID may be used to spatially relate localizer images with a set of axial images, however, in some cases (e.g., multiple localizer images being related to a single set of axial images) a common Frame of Reference UID may not be sufficient. The Referenced Image Sequence (0008,1140) provides an unambiguous method for relating localizer images.	