

## DICOM Change Proposal Form - Final Text

Tracking Information - Administrator Use Only	
Change Proposal Number	CP-26
STATUS	voting-1, approved
Date of Last Update	9-15-94
Person Assigned	Don Van Syckle

<p>Log Summary:            Standard Extended SOP Class should exclude adding Standard Type 1, 1C, 2, 2c attributes, like Specialized SOP Class. (proof review)</p>	
Submitter Name Larry Tarbox	Submission date 9-12-94
Type of Modification Inconsistency	
Name of Document PS 3.2	Version Number
Rationale for change  The description of a Standard Extended SOP Class does not explicitly exclude adding Standard Type 1, 1C, 2, or 2C Attributes. (We mention adding such attributes to a Specialized SOP Class.)	
Sections of document affected  7.1	
Suggested Wording of change  Add the phrase “, or add additional Standard Type 1, 1C, 2, or 2C Attributes” to the end of item c of the first enumerated list (“Standard Extended SOP Classes shall ...”).	