

## DICOM Correction Item

Correction Number	CP-51
Log Summary: Include Curves in CR IODs	
Type of Modification Inconsistency	Name of Standard PS 3.3 - 1993
Rationale for Correction:  Table A.1-1 "Composite information object modules overview" does not show that the curve module may be included in the CR IOD. Table A.2-1 "CR image IOD modules" lists the curve module (C.10.2) as optional module.	
Sections of Document Affected: Table A.1-1	
Correction Wording:  Include the curve module into the CR column of Table A.1-1 as a "U".	