

DICOM Correction Item

Correction Number CP-1012	
Log Summary: Exposure Tag Availability for PORTAL image type	
Type of Modification Addition	Name of Standard PS 3 2009
Rationale for Correction: The following tags are currently excluded for RT Images of type PORTAL, although they can be useful there as well for recording.	
Sections of documents affected PS 3.3	
Correction Wording:	

In PS 3.3, Section C.8.8.2 (RT Image Module), Table C.8-38, change the following text.

Attribute Name	Tag	Type	Attribute Description
Exposure Sequence	(3002,0030)	3	Introduces sequence of Exposure parameter sets, corresponding to exposures used in generating the image. One or more items may be included in this sequence. See C.8.8.2.4.
>X-Ray Tube Current	(0018,1151)	2C	Imaging device X-Ray Tube Current (mA). Required if Value 3 of Image Type (0008,0008) is SIMULATOR or RADIOGRAPH. <u>May be present otherwise.</u>
>Exposure Time	(0018,1150)	2C	Time of X-Ray exposure (msec). Required if Value 3 of Image Type (0008,0008) is SIMULATOR or RADIOGRAPH. <u>May be present otherwise.</u>