

DICOM Correction Item

Correction Number	CP- 1049
Log Summary:	Typo correction following CP-750
Type of Modification	Typo correction, wrong reference table used
Modification	PS 3.3-2009
Rationale for Correction: CP-750 introduced a requesting service code into the core standard. However the table reference used to call the Code sequence macro has erroneously been put to Table 10-1 while in the rest of the Part 3 this macro is in fact to be found in Table 8.8-1. This CP aims at correcting this apparent typo error.	
Sections of documents affected PS 3.3 C.4.12, C.4.18, C.7.2.1	
Correction Wording:	

PS 3.3 C.4.12: Correct table reference for the requesting Service Code Sequence for MWL:

C.4.12 Imaging Service Request Module

**Table C.4-12
IMAGING SERVICE REQUEST MODULE ATTRIBUTES**

Attribute Name	Tag	Attribute Description
...		...
Requesting Service Code Sequence	(0032,1034)	Institutional department where the request initiated. Equivalent to HL7 v2.x ORC-17. Only a single item shall be permitted in this sequence. See Note.
>Include 'Code Sequence Macro' Table 10-1 8.8-1		Defined Context ID is 7030.
...

PS 3.3 C.4.18: Correct table reference for the requesting Service Code Sequence for GP-WL:

C.4.18 General Purpose Scheduled Procedure Step Relationship Module

**Table C.4-18
GENERAL PURPOSE SCHEDULED PROCEDURE STEP RELATIONSHIP MODULE ATTRIBUTES**

Attribute Name	Tag	Attribute Description
...
>Requesting Service Code Sequence	(0032,1034)	Institutional department where the request initiated. Equivalent to HL7 v2.x ORC-17. Only a single item shall be permitted in this sequence. See Note.
>>Include 'Code Sequence Macro' Table 10-1 8.8-1		Defined Context ID is 7030.
...

PS 3.3 C.7.2.1: Correct table reference for the Requesting Service Code Sequence for composite instances:

C.7.2.1 General Study Module

Table C.7-3 specifies the Attributes that describe and identify the Study performed upon the Patient.

Table C.7-3

GENERAL STUDY MODULE ATTRIBUTES

Attribute Name	Tag	Type	Attribute Description
...
Requesting Service Code Sequence	(0032,1034)	3	Institutional department where the request initiated. Only a single item shall be permitted in this sequence.
<i>>Include 'Code Sequence Macro' Table 10-1<u>8.8-1</u></i>		<i>Defined Context ID is 7030.</i>	
...