

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2012/03/27
Person Assigned	Björn Nolte (bjoern.nolte@siemens.com)
Submitter Name	Björn Nolte (bjoern.nolte@siemens.com)
Submission Date	2011/09/23

Correction Number	CP- 1105
Log Summary:	Conformance Claim for SR-CDA Transform
Name of Standard	PS 3.2 - 2011
Rationale for Correction:	Conformance claim regarding DICOM SR to CDA transformation is missing in Part 2.
Correction Wording:	

Add to PS3.2 , “3.11 DICOM Conformance” after 3.11.11 Security Profile

...

3.11.12 Transformation of DICOM SR to CDA

A mechanism for mapping and transforming DICOM SR objects to HL7 CDA documents.

...

Add to PS3.2, “7 Conformance Requirements” after 7.6 SECURITY PROFILES

...

7.7 TRANSFORMATION OF DICOM SR TO CDA

DICOM specifies the transformation of DICOM SR objects to CDA documents in PS 3.20.

This transformation is unidirectional (DICOM SR to HL7 CDA). Conformance statements shall at a minimum state conformance to the top level templates used.

...

Add to PS3.2, "Annex A" after A.5 MEDIA INTERCHANGE

...

A.6 TRANSFORMATION OF DICOM SR TO CDA

The supported SR objects and corresponding template identifiers shall be described. The release version and template identifier of the generated valid CDA documents shall be documented.

...

Increase numbering for the trailing sections of Annex A