

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2011/08/15
Person Assigned	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submitter Name	Ulrich Busch ( <a href="mailto:ulrich.busch@varian.com">ulrich.busch@varian.com</a> )
Submission date	2011/01/18

Correction Number	CP-1118
Log Summary: Recorded Brachy Applicator Terms	
Name of Standard PS 3 2011	
Rationale for Correction There is an inconsistency between (1) RT Plan IOD, RT Brachy Application Setup Module and (2) RT Brachy Treatment Record IOD, RT Brachy Session Record Module: (1) C.8.8.15 RT Brachy Application Setups Module The Source Applicator Type (300A,0292) is specified as "Defined Terms". (2) C.8.8.22 RT Brachy Session Record Module The Source Applicator Type (300A,0292): is specified as "Enumerated Values". This should have been Defined Terms in both places.	
Correction Wording:	

*In PS 3.3, Section C.8.8.22 (RT Brachy Session Record Module), Table C.8-58, correct the following text:*

Attribute Name	Tag	Type	Attribute Description
>>>Source Applicator Type	(300A,0292)	1C	Type of Source Applicator. Required if Recorded Source Applicator Sequence (3008,0140) is sent. <del>Enumerated Values:</del> <b>Defined Terms</b> FLEXIBLE, RIGID.