

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2011/08/15
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Johannes Stahl (Johannes.stahl@siemens.com)
Submission date	2011/01/18

Correction Number	CP-1122
Log Summary: Correct Attribute Tag for Reference Image Sequence	
Name of Standard	PS 3 2011
Rationale for Correction	The attribute description for the Reference Image Number (300A, 00C8) and Referenced Reference Image Number (300C,0007) contains a reference to The Referenced Reference Image Sequence. However the attribute tag is listed incorrectly as "300A,0042" when instead it should be "300C,0042"
Correction Wording:	

In PS 3.3, change the text in the following Section C.8.8.14 (RT Beams Module Attributes)

Attribute Name	Tag	Type	Attribute Description
>>Reference Image Number	(300A,00C8)	1	Uniquely identifies Reference Image within Referenced Reference Image Sequence (300A,0042) (300C,0042) .
>>Referenced Reference Image Number	(300C,0007)	3	Uniquely identifies Reference Image to which planned verification image is related, specified by Reference Image Number (300A,00C8) within Referenced Reference Image Sequence (300A,0042) (300C,0042) .