

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2011/10/17
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Ulrich Busch (ulrich.busch@varian.com)
Submission Date	2011/03/23

Correction Number	CP-1134
Log Summary:	Correction of RT Ion Plan Module Table
Name of Standard	PS 3.3 2011
Rationale for Correction:	In Table A.1-4 COMPOSITE INFORMATION OBJECT MODULES OVERVIEW - RADIOTHERAPY in section A.1.4 Overview of the Composite IOD Module Content. The RT Ion Plan uses the 'RT Ion Tolerance Tables' module, but not the 'RT Tolerance Tables' module as indicated in this table.
Correction Wording:	

In PS 3.3, in Table A.1-4 COMPOSITE INFORMATION OBJECT MODULES OVERVIEW - RADIOTHERAPY in section A.1.4 Overview of the Composite IOD Module Content:

Add a row for 'RT Ion Tolerance Tables' and mark this module to as being used by RT Ion Plan, rather than the currently marked module 'RT Tolerance Tables'.