

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2011/10/17
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Harold Beunk (harold.beunk@nl.nucletron.com)
Submission Date	2011/03/20

Correction Number	CP-1136
Log Summary:	Correct Channel Time Tag Number
Name of Standard	PS 3.3 2011
Rationale for Correction:	There is a wrong tag number used in a text reference.
Correction Wording:	

In PS 3.3 C.8.8.15.6 Time calculations correct the tag number

...

If Treatment Type (300A,0202) is PDR, then the Channel Total Time (300~~8~~A,0286) shall specify the duration of a single pulse.