

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2011/10/25
Person Assigned	Harry Solomon
Submitter Name	Herve Hoehn
Submission date	2011/03/16

Correction Number	CP-1141
Log Summary:	Part 16 consistency clean-up Annex C
Name of Standard	PS 3.16-2011
Rationale for Correction:	<p>Many Template Content Item specifications in Part 16 do not follow the defined formats.</p> <p>This change adds DT or EV to concept name codes, and corrects VT NUM (used in SR templates) to NUMERIC (used in acquisition templates).</p>
Sections of documents affected	PS 3.16 Annex C
Correction Wording:	

**TID 3403**  
**Catheterization Acquisition Context**  
**Type: Extensible      Order: Non-Significant**

	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1	CODE	<u>EV</u> (G-72BB, SRT, "Catheterization Procedure Phase")	1	U		BCID(3250) Catheterization Procedure Phase
2	CODE	<u>EV</u> (109058, DCM, "Contrast Phase")	1	U		BCID(3600) Relative time
3	CODE	<u>EV</u> (109059, DCM, "Physiological challenges")	1	U		BCID(3271) Hemodynamic Physiological Challenges
4	NUMERIC	<u>EV</u> (109060, DCM, "Procedure Step Number")	1	U		UNITS=EV("{step}", UCUM, "step")
5	TEXT	EV (121124, DCM, "Procedure Action ID")	1	U		

**TID 3450**  
**Cardiac Electrophysiology Acquisition Context**  
**Type: Extensible      Order: Non-Significant**

	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1	CODE	<u>EV</u> (109061, DCM, "EP Procedure Phase")	1	U		BCID(3254)
2	<u>NUMERIC</u>	<u>EV</u> (109060, DCM, "Procedure Step Number")	1	U		UNITS=EV("{step}", UCUM, "step")
3	TEXT	<u>EV</u> (109063, DCM, "Pulse train definition")	1	U		

**TID 3470**  
**NM Acquisition Context**  
**Type: Extensible      Order: Non-Significant**

	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1	CODE	<u>DT</u> (109054, DCM, "Patient State")	1	M		DCID (3101) Cardiac Procedural State Values

**TID 8001**  
**Specimen Preparation**  
**Type: Extensible      Order: Significant**

	Value Type	Concept Name	VM	Req Type	Condition	Value Set Constraint
...						
9	INCLUDE	DTID (8003) Specimen Staining	1	MC	IFF Row 3 Processing type value is (P3-00003, SRT, " <u>Staining</u> ")	
...						