

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2012/11/01
Person Assigned	David Clunie dclunie@dclunie.com
Submitter Name	2011/02/28
Submission Date	Kinson Ho kinson.ho@agfa.com on behalf of IHE Radiology Technical Committee

Correction Number	CP-1152
Log Summary: Additional document titles for Key Object Selection Document	
Name of Standard PS 3.16 2011	
<p>Rationale for Correction:</p> <p>IHE Radiology Imaging Object Change Management Profile proposed the use of Key Object Selection Document with specific document titles to support two additional use cases:</p> <ul style="list-style-type: none"> - Correction of incorrect modality worklist selection - Expiration of data retention policy <p>As a result, two new document titles are requested.</p>	
Correction Wording:	

Add two new document titles to CID 7010:

CID 7010 Key Object Selection Document Title

Context ID 7010

Key Object Selection Document Title

Type: Extensible Version: ~~20084028~~20121101

Coding Scheme Designator (0008,0102)	Code Value (0008,0100)	Code Meaning (0008,0104)
...
DCM	113037	Rejected for Patient Safety Reasons
<u>DCM</u>	<u>113038</u>	<u>Incorrect Modality Worklist Entry</u>
<u>DCM</u>	<u>113039</u>	<u>Data Retention Policy Expired</u>

Add definitions for new codes:

113037	Rejected for Patient Safety Reasons	List of objects whose use is potentially harmful to the patient. For example, an improperly labeled image could lead to dangerous surgical decisions.	
--------	-------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------	--

<u>113038</u>	<u>Incorrect Modality Worklist Entry</u>	<u>List of objects that were acquired using an incorrect modality worklist entry, and which should not be used, since they may be incorrectly identified.</u>	
<u>113039</u>	<u>Data Retention Policy Expired</u>	<u>List of objects that have expired according to a defined data retention policy.</u>	