

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2012/01/16
Person Assigned	Harry Solomon
Submitter Name	Harry Solomon (harry.solomon@ge.com)
Submission Date	2011/04/08

Correction Number	CP-1161
Log Summary: Correct Plural Code Meanings	
Name of Standard	PS 3.16 2011
Rationale for Correction	Inconsistencies in code meanings – plural versus singular.
Correction Wording:	

TID 2020 Spectacle Prescription Report Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
8	>	CONTAINS	TEXT	EV (121106, DCM, "Comments")	1	U		

TID 2101 Macular Grid Thickness and Volume Measurement Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
27	>	CONTAINS	TEXT	EV (121106, DCM, "Comments")	1	U		

TID 3202 VENTRICULAR ANALYSIS Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
10	>>	CONTAINS	IMAGE	EV (121112, DCM, "Source of Measurements")	1-n	M		
...								

TID 3205
CALIBRATION
Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
10	>	CONTAINS	INCLUDE	DTID (300) Measurement	1	M		\$Measurement = EV (111066, DCM, "Vertical Pixel Spacing") \$Unit = DT (mm/{pixel}, UCUM, "mm/pixel") \$imagePurpose = EV (121112, DCM, "Source of Measurements")
...								

TID 3208
FRAME-TO-FRAME RESULT
Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
3	>	CONTAINS	IMAGE	EV (121112, DCM, "Source of Measurements")	1-2	M	VM = 1: Single plane analysis, VM = 2: Biplane analysis	
...								

TID 3214
ANALYZED SEGMENT
Type: Extensible

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
3	>	CONTAINS	IMAGE	EV (121112, DCM, "Source of Measurements")	1	M		
...								

TID 3515
Cardiac Output Measurement by Indicator Dilution
Type: Extensible Order: Significant

	NL	Relation with Parent	Value Type	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
2	>	CONTAINS	INCLUDE	TID (300) Measurement	1	M		\$Measurement = EV (8737-9, LN, "Cardiac Output by Indicator Dilution") \$Units = EV (l/min, UCUM, "l/min") \$Method = DCID (3628) Cardiac Output Methods

	NL	Relation with Parent	Value Type	Concept Name	VM	Req Type	Condition	Value Set Constraint
								\$WavePurpose = DT (121112, DCM, "Source of m Measurement")
...								

TID 3530
Hemodynamic Acquisition Context
Type: Extensible Order: Significant

	NL	Relation with Parent	Value Type	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
4		HAS ACQ CONTEXT	WAVEFORM	DT (121112, DCM, "Source of m Measurement")	1	UC	XOR Row 5	
5		HAS ACQ CONTEXT	TCOORD	DT (121112, DCM, "Source of m Measurement")	1	UC	XOR Row 4	
...								

TID 3715
ECG Measurement Source
Type: Extensible Order: Significant

	NL	Relation with Parent	Value Type	Concept Name	VM	Req Type	Condition	Value Set Constraint
...								
3			TCOORD	EV (121112, DCM, "Source of m Measurement")	1	U		
...								

Annex D DICOM Controlled Terminology Definitions (Normative)

...

DICOM Code Definitions (Coding Scheme Designator "DCM" Coding Scheme Version "01")

Code Value	Code Meaning	Definition	Notes
...			
121112	Source of m Measurement	Image or waveform used as source for measurement	
...			