

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2012/01/17
Person Assigned	David Clunie dclunie@dclunie.com
Submitter Name	David Clunie dclunie@dclunie.com
Submission Date	2011/05/22

Correction Number	CP-1166
Log Summary: Allow Frame of Reference in Segmentation even if derived image references	
Name of Standard PS 3.3 2011	
Rationale for Correction: There is no reason to prevent a Frame of Reference UID being sent in a raster Segmentation instance even if the images from which it is derived is explicitly specified, and it is helpful to be able to specify, for example, a common frame of reference explicit. This correction is not applicable to the Surface Segmentation, since the Frame of Reference Module is already mandatory in that.	
Correction Wording:	

Modify Table A.51-1

A.51.3 Segmentation IOD Module Table

**Table A.51-1
SEGMENTATION IOD MODULES**

IE	Module	Reference	Usage
...
Frame of Reference	Frame of Reference	C.7.4.1	C - Required if Derivation Image Functional Group (C.7.16.2.6) is not present. <u>May be present otherwise.</u>
...