

Correction Number		CP-116
Log Summary: Corrections in Modality Worklist Attributes		
Type of Modification: Corrections	Name of Standard PS 3.4-1996	
Rationale for Correction:		
<p>There are two inconsistencies in Table K.6-1 with the attributes for the Modality Worklist:</p> <ol style="list-style-type: none"> <li>1. Tag (0040,0005) is the wrong tag for the Scheduled Procedure Step Description</li> <li>2. In the section with the <b>Imaging Service Request</b> attributes the last cell says: All other attributes from Scheduled Procedure Step Module.</li> </ol>		
Sections of document affected:		
Table K.6-1		
Correction Wording:		
<ol style="list-style-type: none"> <li>1. Replace the tag (0040,0005) with tag (0040,0007):  Scheduled Procedure Step Description           (0040,0007)</li> <li>2. Replace the text in the Imaging Service Request section:  All other Attributes from the Imaging Service Request Module</li> </ol>		