

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2012/03/22
Person Assigned	Bas Revet (bas.revet@philips.com)
Submitter Name	Renate Höcker Renate.Hoecker@siemens.com
Submission Date	2011/06/21

Correction Number	CP-1187
Log Summary:	Wrong DICOM Tags are referenced in part 2
Name of Standard	PS 3.2 - 2011
Rationale for Correction:	Wrong DICOM Tags are assigned in Table C.8.1-2.
Correction Wording:	PS 3.2, C.8.1-2

Update PS 3.2, C.8.1-2 as follows:

**Table C.8.1-2
HL7/MODALITY WORKLIST ATTRIBUTE MAPPING**

DICOM Attribute	DICOM Tag	HL7 Attribute Name	HL7 Segment	Notes
Scheduled Procedure Step				
Scheduled Procedure Step Sequence	(0040,0100)			
> Scheduled Station AET	(0040,0001 2)			DICOMRis generated
> Scheduled Procedure Step Start Date	(0040,0002 3)	Quantity/Timing	ORM OBR:27	DICOMRis generated
> Scheduled Procedure Step Start Time	(0040,0003 6)	Quantity/Timing	ORM OBR:27	DICOMRis generated
> Modality	(0008,0060)			DICOMRis generated
.....				