

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2013/01/29
Person Assigned	Bas Revet ( <a href="mailto:bas.revet@philips.com">bas.revet@philips.com</a> )
Submitter Name	Dave Harvey <a href="mailto:dave@medicalconnections.co.uk">dave@medicalconnections.co.uk</a>
Submission Date	2012/04/24

Correction Number	CP-1233
Log Summary: Add Patient Size (Height) to Modality Worklist	
Name of Standard PS 3.4 2011	
Rationale for Correction: Patient Height & Weight are both useful fields for dose calculations, but only weight is included explicitly in the list of Modality Worklist attributes, leading to developers and users believing that it cannot be used. The following change does not change the semantics or requirements, since the Size remains Type 3, as it has been as part of the Patient Demographic Module, but it makes it explicit that it is allowed. Note that "Size" is defined for humans to be the height in meters.	
Correction Wording:	

*Modify Table K.6-1 as follows*

Patient's Weight	(0010,1030)	○	2	
<b>Patient's Size</b>	<b>(0010,1020)</b>	<b>○</b>	<b>3</b>	
Confidentiality constraint on patient data	(0040,3001)	○	2	
All other Attributes from the Patient Demographic Module		○	3	