

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2013/06/13
Person Assigned	Andrei Leontiev (andrei.leontiev@ge.com)
Submitter Name	Hervé Hoehn, Harry Solomon (harry.solomon@ge.com)
Submission Date	2012/06/16

Correction Number	CP-1272
Log Summary: DCID for Referenced Frame	
Name of Standard PS 3.3 2011	
Rationale for Correction Supplement 145 for Whole Slide Imaging specified a Context Group for the Purpose of Reference in the Referenced Image Functional Group, appropriate to the use case for WSI. Technically this conflicts with the DCID specified in Table C.7.6.16-6	
Correction Wording:	

Amend Section A.32.8.4.1

A.32.8.4.1 VL Whole Slide Microscopy Functional Group Macros Content Constraints

A.32.8.4.1.1 Referenced Image

Baseline Defined Context ID for Purpose of Reference Code Sequence (0040,A170) **in the Referenced Image Functional Group** is CID 8120.

Amend Section C.7.6.16.2.5

C.7.6.16.2.5 Referenced Image Macro

Table C.7.6.16-6 specifies the attributes of the Referenced Image Functional Group macro.

**Table C.7.6.16-6
REFERENCED IMAGE MACRO ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
Referenced Image Sequence	(0008,1140)	2	The set of images or other composite SOP Instances used to plan the acquisition, if any, and other significant related images. See Section C.7.6.16.2.5.1 for further explanation. Zero or more Items shall be included in this Sequence.
<i>>Include 'Image SOP Instance Reference Macro' Table 10-3</i>			
>Purpose of Reference Code Sequence	(0040,A170)	1	Describes the purpose for which the reference is made. Only a single Item shall be included in this sequence. See C.7.6.16.2.5.1 for further explanation.
<i>>>Include 'Code Sequence Macro' Table 8.8-1</i>			<i>Defined CID 7201, <u>or as specified in the IOD invocation of this Functional Group.</u></i>