

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2013/06/13
Person Assigned	Andrei Leontiev (andrei.leontiev@ge.com)
Submitter Name	Hervé Hoehn, Harry Solomon (harry.solomon@ge.com)
Submission Date	2012/06/16

Correction Number	CP-1273
Log Summary: Navigation Sequence Item references one frame	
Name of Standard	PS 3.3 2011
Rationale for Correction	The Referenced Image Navigation Sequence, used in Whole Slide Microscopy LOCALIZER images, should allow only a single frame to be referenced in each Item. PS3.6 Data Dictionary allows this to be VM=1-n
Correction Wording:	

Amend Section C.8.12.7

C.8.12.7 Multi-Resolution Navigation Module

Table C.8.12.7-1 specifies the Attributes that describe the Multi-Resolution Navigation Module, used in Whole Slide Microscopy LOCALIZER images.

**Table C.8.12.7-1
MULTI-RESOLUTION NAVIGATION MODULE ATTRIBUTES**

Attribute Name	Tag	Type	Attribute Description
Referenced Image Navigation Sequence	(0048,0200)	1	Identification of correspondence between areas of the Pixel Data (07FE,0010) and frames of higher resolution images. One or more Items shall be included in this sequence.
>Include 'SOP Instance Reference Macro' Table 10-11			Referenced Image Instance
>Referenced Frame Number	(0008,1160)	1	Referenced Frame Number within Referenced Image. <u>Only a single value shall be present.</u>
...			