

DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2013/08/09
Person Assigned	Ulrich Busch (ulrich.busch@varian.com)
Submitter Name	Ulrich Busch (ulrich.busch@varian.com)
Submission Date	2013/02/07

Correction Number	CP-1293
Log Summary: Remove Invocation Statement For Scheduled Parameters	
Name of Standard PS 3 2011 Part 16	
<p>Rationale for Correction:</p> <p>CP1199 introduced the Content Item Modifier Sequence from Protocol Context Sequence to Scheduled Processing Parameters Sequence in the Unified Procedure Step Scheduled Procedure Information Module and the Unified Procedure Step Performed Procedure Information Module.</p> <p>While doing that, a text block telling 'Context ID may be defined in the macro invocation' was added as well. However, it is no possible to specify CIDs in an Macro invocation, since the macro is invocated in a module already end therefore there is no facility to do what is mentioned.</p> <p>This CP therefore replaces those clauses by 'No Baseline CID defined'.</p>	
Correction Wording:	

In PS 3.3, make the following modifications:

C.30.2 Unified Procedure Step Scheduled Procedure Information Module

Table C.30.2-1 specifies the Attributes that describe the Unified Procedure Step (UPS) to be performed. The UPS may or may not be scheduled for a specific time or device, or may simply represent a piece of work that is intended to be performed.

**Table C.30.2-1
Unified Procedure Step Scheduled Procedure Information Module Attributes**

Attribute Name	Tag	Attribute Description
...
Scheduled Workitem Code Sequence	(0040,4018)	Coded description of the Procedure Step. Only a single Item shall be permitted in this sequence.
<i>>Include Code Sequence Macro Table 8.8-1</i>		<i>Baseline Context ID is CID 9231</i> <i>Note: This CID has generic workitems. An implementation may choose to define more specific, detailed workitems.</i>
Scheduled Processing Parameters Sequence	(0074,1210)	Processing parameters to be used by the performing system when carrying out the Procedure Step.

Attribute Name	Tag	Attribute Description
		Zero or more Items shall be included in this sequence.
>Include Content Item Macro Table 10-2		<p><u>No Baseline Context IDs are defined for Concept Name Code Sequence or Concept Code Sequence.</u></p> <p><u>Context ID may be defined in the macro invocation</u></p>
>Content Item Modifier Sequence	(0040,0441)	Sequence that specifies modifiers for a Protocol Context Content Item. One or more Items are permitted in this sequence.
>> Include 'Content Item Macro' Table 10-2		<p><u>No Baseline Context IDs are defined for Concept Name Code Sequence or Concept Code Sequence.</u></p> <p><u>Context ID may be defined in the macro invocation</u></p>
Input Readiness State	(0040,4041)	<p>Readiness state of the Input Information Sequence (0040,4021) and the referenced instances.</p> <p>Enumerated values are:</p> <p style="padding-left: 40px;">INCOMPLETE</p> <p style="padding-left: 40px;">UNAVAILABLE</p> <p style="padding-left: 40px;">READY</p> <p>INCOMPLETE means the Input Information Sequence is not yet complete and additional instance references might be added.</p> <p>UNAVAILABLE means the Input Information Sequence is complete but one or more of the referenced instances might not yet be available from the referenced source(s).</p> <p>READY means that the Input Information Sequence is complete and the referenced instances are available from the referenced sources.</p> <p>Notes:</p> <ol style="list-style-type: none"> 1) If the Procedure Step does not require input information, the Input Readiness State may be READY when the Input Information Sequence is empty. 2) There is no guarantee that the referenced instances will still be available at the referenced location when retrieval is attempted. 3) There is no requirement to confirm the presence of referenced media prior to setting the Input Readiness State to READY.
...

...

C.30.3 Unified Procedure Step Performed Procedure Information Module

Table C.30.3-1 specifies the Attributes that describe the performance and results of a Unified Procedure Step (UPS).

**Table C.30.3-1
Unified Procedure Step Performed Procedure Information Module Attributes**

Attribute Name	Tag	Attribute Description
...
>Performed Workitem Code Sequence	(0040,4019)	A sequence that conveys the type of procedure performed. Zero or more items shall be present in this sequence.
>>Include Code Sequence Macro Table 8.8-1		<i>Baseline Context ID is CID 9231</i> Note: This CID has generic workitems. An implementation may choose to define more specific, detailed workitems.
>Performed Processing Parameters Sequence	(0074,1212)	Parameters used to perform the procedure. Zero or more items shall be present in this sequence
>>Include Content Item Macro Table 10-2		<i>No Baseline Context IDs are defined for Concept Name Code Sequence or Concept Code Sequence.</i> <i>Context ID may be defined in the macro invocation</i>
>>Content Item Modifier Sequence	(0040,0441)	Sequence that specifies modifiers for a Protocol Context Content Item. One or more Items are permitted in this sequence.
>>>Include 'Content Item Macro' Table 10-2		<i>No Baseline Context IDs are defined for Concept Name Code Sequence or Concept Code Sequence.</i> <i>Context ID may be defined in the macro invocation</i>
>Performed Procedure Step End DateTime	(0040,4051)	Date and Time at which the Procedure Step ended.