

## DICOM Correction Proposal

STATUS	Final Text
Date of Last Update	2013/10/10
Person Assigned	David Clunie <a href="mailto:dclunie@dclunie.com">dclunie@dclunie.com</a>
Submitter Name	David Clunie <a href="mailto:dclunie@dclunie.com">dclunie@dclunie.com</a>
Submission Date	2013/03/16

Correction Number	CP-1298
Log Summary: Update description of Detector ID de-identification	
Name of Standard	PS 3.15 2011
Rationale for Correction:	The Breast Tomosynthesis Contributing Sources Module contains Detector ID as Type 1, so the Basic Application Level Confidentiality Profile in PS 3.15 should be updated from X (remove) to X/D (remove or replace with dummy value depending on the IOD).
Correction Wording:	

Amend PS 3.15:

**Table E.1-1  
Application Level Confidentiality Profile Attributes**

Attribute Name	Tag	Retired (from PS 3.6)	In Std. Comp. IOD (from PS 3.3)	Basic Profile	Retain Safe Private Option	Retain UIDs Option	Retain Device Ident. Option	Retain Patient Chars. Option	Retain Long. Full Dates Option	Retain Long. Modif. Dates Option	Clean Desc. Option	Clean Struct. Cont. Option	Clean Graph. Option
...	...	...	...	...									
Detector ID	(0018,700A)	N	Y	<u>X/D</u>			K						
Device Serial Number	(0018,1000)	N	Y	X/Z/D			K						
...	...	...	...	...									