

Correction Number		CP-138	
Log Summary: Procedure and Action Item Code Sequence for Image IODs			
Type of Modification		Name of Standard	
Addition		PS 3.3-1998	
<p>Rationale for Correction</p> <p>There is no place in the existing Composite Image objects for recording the Procedure Code or Action Items actually performed.</p> <p>What was scheduled via the modality worklist can be recorded in the Request Attributes Sequence in the image, but this may not correspond to what was actually performed.</p> <p>The appropriate attributes that could also be recorded in the image are defined in modules of the Modality Performed Procedure Step IOD</p> <p>The Procedure Code Sequence attribute should be placed at the Study level, since it corresponds to the request, but it may have multiple values since the image may be created to satisfy more than one request.</p> <p>The Performed Action Item Code Sequence should be placed at the Series level, since there is no Performed Procedure Step/Study Component entity explicitly represented in a Composite IOD, and the Series level is the next entity lower.</p> <p>This is consistent with the approach of Supplement 17 that added Modality Performed Procedure Step.</p> <p>One of the potential uses of these codes is to allow devices receiving images to route or hang (organize display layout) images based on codes. At the present time, other text fields are used by different implementors in an ad hoc manner for this purpose. Even in the absence of a single universal coding scheme for procedures and action items having a consistent place to enter common codes is helpful.</p>			
Sections of documents affected			
C.7.2.1 General Study Module Table C.7-2			
C.7.3.1 General Series Module Table C.7-4			
Correction Wording:			

Append to C.7.2.1 General Study Module Table C.7-2:

C.7.2.1 General Study Module

...

**Table C.7-2
 GENERAL STUDY MODULE ATTRIBUTES**

...
Procedure Code Sequence	(0008,1032)	3	A Sequence that conveys the type of procedure performed. One or more Items may be included in this Sequence.
<i>>Include 'Code Sequence Macro' Table 8.8-1</i>		<i>No Baseline Context ID is defined.</i>	

Append to C.7.3.1 General Series Module Table C.7-4:

C.7.3.1 General Series Module

...

**Table C.7-4
GENERAL SERIES MODULE ATTRIBUTES**

...
Performed Action Item Sequence	(0040,0260)	3	Sequence describing the Action Items performed for this Procedure Step. One or more Items may be included in this Sequence.
<i>>Include 'Code Sequence Macro' Table 8.8-1</i>		<i>No Baseline Context ID is defined.</i>	